



Phasemeter User Manual

Moku:Lab's Phasemeter measures phase with better than $6 \mu\text{radian}$ precision for input signals oscillating between 1 kHz and 200 MHz. Using a digitally implemented phase-locked loop architecture, it provides exceptional dynamic range and precision far exceeding the capabilities of conventional lock-in amplifiers and frequency counters. The Moku:Lab's Phasemeter is ideal for applications demanding precise measurements of phase or frequency, including precision metrology and heterodyne interferometry, channel characterisation in communication networks, clock recovery and signal reconditioning for digital communication systems, and laser frequency stabilization.





Table of Contents

Introduction	4
How does it work?	4
Why use a Phasemeter?	4
Heterodyne interferometry	4
User Interface.....	5
Main Menu	6
Channel Configuration.....	7
Channels	7
Acquisition frequency	8
Bandwidth	8
Input voltage range	8
Acquisition speed	8
Freewheeling	8
Outputs	9
Phase-locked output	9
Measurement Data	10
Measurement tabs	10
Frequency	10
Phase	10
Amplitude	10
Lock status indicator	10
Reacquisition	10
Data Visualization.....	11
Plot Types	12
Timeseries	12
Power spectral density	13
Amplitude spectral density	14
Coherence	15
Allan deviation	16
Data Acquisition.....	17
Exporting Data	18
Live Data	18
Logged data	19
SD card	19
Example Measurement Configurations	20
Measure the relative phase of two signals	20



Ensure Moku:Lab is fully updated. For the latest information:


www.liquidinstruments.com



Introduction

How does it work?

Moku:Lab's Phasemeter measures phase using a digitally implemented phase-locked loop, which uses feedback control to continuously update the phase of a local oscillator to equal that of the input signal.



The instantaneous phase error between the input signal and local oscillator is detected via demodulation using a digital multiplier and low-pass filter, almost identical in principle to a lock-in amplifier. The detected phase error is then passed through a PID controller to generate a feedback control signal to continuously update the phase of the local oscillator. The phase of the input signal relative to the local oscillator is measured by keeping a record of every change made to the phase of the local oscillator.

Why use a Phasemeter?

Digitally implemented phase-locked loops have extremely high dynamic range, allowing them to contiguously measure phase over millions of cycles with a sensitivity of better than 6 micro-radians. This is particularly important for applications where phase that expected to drift over many wavelengths within the measurement time, but still require extremely high measurement precision.

Heterodyne interferometry


One key application of phasemeters is in heterodyne interferometry, where displacement information is stored within the phase of a beat-note produced by interfering two electric fields with slightly different frequencies at a photodetector. Laser heterodyne interferometers are typically used to measure tiny displacements on the order of a fraction of the laser wavelength.

At a laser wavelength of 1064 nano-meters, Moku:Lab's phasemeter is capable of measuring displacements with pico-meter sensitivity (i.e., *one millionth* of the wavelength of the laser). It is not uncommon, however, for heterodyne interferometers to experience displacements on the order of many thousands of wavelengths due to path-length contraction and expansion caused by shock, vibrations and changes in temperature. And depending on their optical configuration, heterodyne interferometers can also be extremely susceptible to laser frequency noise which typically appears as large, random excursions in phase at low frequencies.

The ability to measure phase with high dynamic range is therefore crucial in heterodyne interferometry.



User Interface




ID	Description	ID	Description
1	Main menu	8	Output settings
2	Display frequency data	9	Data logger
3	Display phase data	10	Start / pause measurement
4	Export data	11	Data visualization
5	Display amplitude data	12	Channel data display area 1
6	Instrument configuration menu	13	Reacquire button
7	Channel settings	14	Channel data display area 2



Main Menu

The **main menu** can be accessed by pressing the  icon, allowing you to:






Channel Configuration

The **channel configuration menu** allows you to configure the Phasemeter's channel settings and outputs.

Access the measurement configuration menu by pressing the  icon.

Channels

● Show / hide the measurement configuration menu



The screenshot shows the 'Channels' configuration menu. At the top, there are icons for 'Channels' and 'Output'. Below this, the 'Input 1' section is active, indicated by a red toggle switch. The settings for Input 1 are: Frequency (Auto, 1.345 812 143 280 MHz), Bandwidth (150 Hz), Coupling (AC), Impedance (50 Ω), and Range (1 V_{pp}, 10 V_{pp}). The 'Input 2' section is inactive, with a grey toggle switch. Below the input sections is the 'Acquisition speed' section, which has a grid of buttons for different speeds: 30.5 Hz, 122 Hz, 488 Hz, 1.95 kHz, 15.6 kHz, and 125 kHz. At the bottom is the 'Freewheeling' section, which is currently enabled (blue toggle switch). A vertical waveform display is visible on the left side of the screen.

- ← Activate / deactivate channel 1
- ← Enable **auto-acquisition** mode
- ← Configure the **acquisition frequency** of the phasemeter
- ← Select the **bandwidth** of the phasemeter
- ← Select between **AC** and **DC** coupling
- ← Select between **50 Ω** and **1 MΩ** input impedance
- ← Select between **1 V_{pp}** and **10 V_{pp}** input range
- ← Select the data **acquisition speed**
- ← Enable / disable **freewheeling**



Acquisition frequency

- The phasemeter will attempt to track frequencies around the specified acquisition frequency
- If you know the frequency of the tone you'd like to measure, you can specify it manually by tapping the blue number below the Frequency label
- If you do not know the frequency of the tone you'd like to measure, you can enable auto-acquisition mode. This will automatically search for and track the highest magnitude tone between 500 kHz and 200 MHz

Note: Auto-acquisition does not work reliably for tones below 500 kHz

Bandwidth

- The Moku:Lab's Phasemeter will reliably measure the phase of an input signal up to the specified bandwidth
- Select between 10 Hz, 40 Hz, 150 Hz, 600 Hz, 2.5 kHz and 10 kHz bandwidth settings
- **Note:** The selected bandwidth should not exceed one fifth of the acquisition frequency

Input voltage range

- Select an appropriate input voltage range to avoid harmonic distortion caused by clipping
- Input sensitivity is 10 times lower at 10 V_{pp} input voltage range. If the amplitude of the input signal is lower than 1 V_{pp}, use the 1 V_{pp} input voltage range setting

Acquisition speed

- Acquisition speed specifies the sampling rate at which phase, frequency and amplitude data is saved to file or streamed over a network.
- Data visualization (graphs) are not available for acquisition speeds above 500 Hz

Freewheeling


- When freewheeling mode is enabled, the phasemeter will continue to 'freewheel' at a constant frequency when the input signal is too weak to track reliably. The phasemeter will resume tracking the phase of the input signal when its amplitude returns to a reliable level
- The freewheeling frequency is determined by averaging the previous two seconds of continuous, uninterrupted frequency data
- Freewheeling mode is useful in applications where the amplitude of the input signal is expected to fluctuate significantly. For example, freewheeling is useful in free-space optical communications systems where the phasemeter can be used to perform clock recovery in the presence of strong atmospheric turbulence



Outputs

The Phasemeter features two output sine generators with manual control over amplitude, frequency and phase. The outputs can also be phase-locked to their corresponding input channel whilst maintaining the full range of control over amplitude and phase.

The phase of the two outputs can be synchronized by tapping the **Sync phase** button at the bottom of the tab.




Phase-locked output

- Generate an output tone with the same frequency and phase of the input signal
- The amplitude and phase of the generated tone remains configurable



Measurement Data



Measurement tabs

Frequency

- The Frequency measurement tab displays the input signal's frequency in Hertz (Hz)

Phase

- The Phase measurement tab displays the input signal's phase in units of cycles (cyc), radians (rad) or degrees (deg)
- Tap the blue 'units' text to switch between units
- Zero the phase offset by tapping the icon on the right-hand side of the display. Zeroing the phase offset of the Delta channel will zero the phase offset of channels 1 and 2

Amplitude

- The Amplitude measurement tab displays the input signal's amplitude in units of Volts RMS (V_{rms}), Volts peak-to-peak (V_{pp}) or Decibels (dB)
- Tap the blue 'units' text switch between units

Lock status indicator

- Indicates whether or not the phasemeter is tracking the input signal correctly. A red icon indicates that the phasemeter is not tracking the input signal


Reacquisition

- Tap the **Reacquire** button to reset both phasemeter channels simultaneously
- Both channels are reset at the same time to maintain synchronization



Data Visualization

The data visualization panel can be accessed by tapping the  icon at the bottom left of the interface, allowing you to display measurement data in a variety of formats and over different time and frequency scales.






Plot Types

Frequency, Phase and Amplitude data can be displayed in different formats, including timeseries, power spectral density, amplitude spectral density, coherence, Raleigh spectrum and Allan deviation.

- All plot types can be auto-scaled by double tapping anywhere on the graph
- Individual traces can be hidden and revealed by tapping the **1**, **2** and **Δ** icons located at the top left of the graph

Timeseries


- Timeseries data can be viewed over time spans ranging from 0.5 seconds to 600 seconds
- Adjust timescale and span using pinch gestures anywhere on the graph
- Set the start and end times of the span manually using the slide rule located above the graph





Power spectral density


- Power spectral density describes a signal's distribution of power at different frequencies.
- The units of power spectral density are proportional to $\text{amplitude}^2/\text{Hz}$ (e.g., $\text{cycles}^2/\text{Hz}$)





Amplitude spectral density


- Amplitude spectral density provides a measure of a signal's amplitude at different frequencies
- The units of amplitude spectral density are proportional to amplitude/ $\sqrt{\text{Hz}}$ (e.g. cycles/ $\sqrt{\text{Hz}}$)
- Amplitude spectral density is equal to the square root of the power spectral density





Coherence

- Spectral coherence is a unitless statistic used to measure the similarity between two signals





Allan deviation

- Allan deviation is a unitless measure of stability, typically used to quantify the stability of clocks and other oscillators
- Allan deviation is equal to the square-root of the Allan variance
- An Allan deviation of 2×10^{-6} at an averaging time of $\tau = 1$ seconds can be interpreted as there being an RMS error between two measurements one second apart of 2×10^{-6} cycles






Data Acquisition

The Moku:Lab's Phasemeter can acquire data at a maximum **acquisition speed** of 62.5 kS/s for two channels and 125 kS/s for one channel. To access the data acquisition menu, press  icon.

- Data can be saved to SD card or RAM with binary ***.li** or comma separated value ***.csv** file formats
- Files saved to RAM will be lost when the Moku:Lab is powered down or reset
- Files saved with binary ***.li** format can be converted to ***.csv** or ***.mat** using Liquid Instruments file conversion software (<https://github.com/liquidinstruments/lireader>)
- Record data for up to 240 hours, and delay the start of a measurement for up to 240 hours
- Start a measurement by pressing the red circle



Note: As a precaution, you will be warned about switching instruments while a measurement is taking place.




Exporting Data

Export data by pressing the  icon.

Live Data

Measurement traces can be uploaded to My Files (iOS 11 or later), Dropbox, E-mail, SD card, iCloud, Clipboard (screenshot is not copied to the clipboard).

To export a live data, tap the  icon and select the 'Live Data' option.






Logged data

Data that has been acquired to SD card or RAM can be uploaded to My Files (iOS 11 or later), Dropbox, E-mail, and iCloud.

To export logged data, press the  icon and select the 'Logged data' option.



SD card

- Upload files to SD card by inserting a compatible FAT32 formatted drive into the Moku:Lab's SD card slot, located on the rear of the device next to the power connector.



Example Measurement Configurations

Measure the relative phase of two signals

To measure the phase of one signal with respect to another:

1. Connect the two signals to the Moku:Lab's inputs.
2. Configure the two input channels for your measurement setup.
 - a. The Acquisition speed and Bandwidth settings limit the range of frequencies within which you can observe changes in magnitude and phase. For example, to observe features up to 200 Hz, set the Bandwidth to be at least 600 Hz and the acquisition rate to be at least 488 Hz.

Note: When measuring the relative frequency, phase and amplitude of two signals, it's often useful to configure both channels identically to maximise the rejection of common sources of error and noise in the Delta measurement.

3. When both channels have been configured, tap the **Reacquire** button to synchronously reset both phasemeter channels.
4. View the data in the frequency and time domains by tapping the  icon. Double-tap the graph to automatically scale the vertical axis, and adjust the horizontal axis using the slider located above the graph or by using pinch gestures.

Tip: Tap the 'Clear' button at the top right of the graph every time you reacquire to discard transient data which can sometimes corrupt the quality of the graph

5. To record data, tap the  icon and configure the data logger as required for the measurement.

Note: If the Moku:Lab's internal clock is not synchronized to that of the device generating the input signals, you can expect the measured phase for channels 1 and 2 to 'ramp' linearly over time.

The reason this occurs is because phase is the integral of frequency, which means that any DC frequency error between the Moku:Lab's internal clock and that of the external source will cause the measured phase to grow at a rate proportional to the frequency difference between the two devices.

As long as the two input signals are generated by the same source, the frequency error will be common to both phase measurements and will be cancelled out in the Delta phase measurement.



Ensure Moku:Lab is fully updated. For the latest information:

www.liquidinstruments.com