



**ST102R. Giardia-a-Glo™ Reagent-Only Kit**  
 Fluorescein-labeled Monoclonal Antibody Reagent for  
 Direct Immunofluorescence Detection of *Giardia* Cysts in Stool Specimens

---

**Explanation** The Giardia-a-Glo™ kit is intended for use in diagnosis of human or animal infection with the parasites, *Giardia lamblia*, by direct immunofluorescence. The test is designed for testing thin smears of stool (fecal) specimens made on our specially treated multi-well microscope slides. The test demands access to and skill in using a fluorescence microscope.

**Description of Products**

- » *Giardia lamblia* is a common, ubiquitous intestinal parasitic protozoan that causes gastroenteritis in man and lower animals. It is spread through food, water, and fomites. A range of reservoir host animals exists for *Giardia*. Symptoms of infection may include diarrhea, intestinal cramping, bloating, gas, vomiting, and, occasionally, low-grade fever. The cellular stages that would appear in the feces of infected persons or lower animals would include the oval-shaped cyst and, far less commonly, the pear-shaped trophozoite. The *Giardia* cyst measures approximately 8-13 um in length and 7-10 um in width; the trophozoite, approximately 9-21 um long by 5-15 um wide.
- » The Giardia-a-Glo™ kit is designed to detect the cyst stage of this parasite and will not detect the *Giardia* trophozoite.
- » The Giardia-a-Glo™ kit utilizes the principle of direct immunofluorescence. The antibody reagent of a fluorescein-labeled mouse monoclonal antibody made to a cyst wall antigenic site (epitope) of *Giardia lamblia*. The reagent will bind only to the cysts of this parasite if they are present in the stool or feces. The cysts will appear bright apple-green when viewed under a fluorescence microscope using the appropriate filters for fluorescein.
- » Reagent included is a working dilution (1x) of a fluorescein (FL)-labeled monoclonal antibody made against cysts of *Giardia lamblia* (= *G. intestinalis*). This kit provides enough reagent for at least 75 tests, using one drop per test (approximately 45 microliters per drop). The antibody reagent contains 0.04% w/v sodium azide as preservative and 1% bovine serum albumen as antibody stabilizer. This reagent shows varying degrees of cross-reactivity with cysts of other species of *Giardia*.
- » Positive Control is a mixture of *Giardia lamblia* cysts and *Cryptosporidium parvum* oocysts in a buffered formaldehyde solution with fecal material. The concentration of this suspension is approximately 2x10e5 cysts and oocysts (each) per mL.

**Storage: Store at 4° C. DO NOT FREEZE.**  
**ST102R reagent is light sensitive.**

**Kit Includes:**

- ST102R: 1 dropper vial containing 3.5 mL working dilution (1x) reagent
- PC101: 1 screw cap vial containing 1.0 mL positive control

**Other Lab Supplies Not Included, but Available:**

- B100-40: 40 mL Dilution/Blocking (D/B) Buffer
- C101: 3.5 mL BlockOut™ counterstain
- D101: 0.4 mL DAPI, 5000X in methanol
- M101: 3.5 mL No-Fade™ Mounting Medium
- M102: 3.5 mL Elvanol No-Fade™ Mounting Medium
- S100-1-9MM: One-well (9mm) SuperStick™ Slides, 40/box
- S100-1: One-well (14mm) SuperStick™ Slides, 40/box
- S100-2: Two-well SuperStick™ Slides, 40/box
- S100-3: Three-well SuperStick™ Slides, 40/box
- WB100: 1 screw cap bottle containing 50 ml 20x SureRinse™ Wash Buffer
- WB101: 1 screw cap bottle containing 50 mL 1x SureRinse™ Wash Buffer
- PACIR: AccuSpike™ -IR, G/C Quality Control Standard (PACIR3, PACIR6, PACIR12)

**Preparation**

1. Prepare environmental sample(s) to be applied to well slide.
- 

Contact us by email for MSDS or Certificate of Analysis/QC Report.  
 Email: [contact@waterborneinc.com](mailto:contact@waterborneinc.com)

## Instructions for Use

1. Isolated water particulates should be air-dried onto a well of a pre-treated slide, using a stream of warm (not hot) air; alternatively, a slide-warmer may be used. Do not allow the slide to become hot to the touch. Samples must be completely dry before continuing to step 2. (Drying time: Approximately 15 – 30 minutes.)
2. A methanol fixation step **may** be performed at this point, however, **it is not required for this reagent to bind well to cysts and oocysts**. Methanol fixation may intensify DAPI staining. Methanol fixation: Apply 45-uL absolute methanol to the well of the slide. Allow the well of the slide to dry completely. (Drying time: Approximately 30 minutes.)
3. When the sample has dried completely, DAPI staining may be performed here. Add 50 uL of a working dilution (1X) of 4',6-diamidino-2-phenylindole (DAPI) to each sample well. Leave on sample for 1 minute at room temperature.
4. Rinse the slide free of DAPI by adding 50 – 100 uL SureRinse™ wash buffer and leave for 1 minute. Tilt slide, long edge down, and absorb excess fluid with absorbent material placed at the edge of the slide well. Do not touch the surface of the well slide or disturb the sample.
5. Apply one drop (approximately 45 uL) of Giardia-a-Glo™ antibody reagent to the spot of dried test particulates in each well. If necessary, spread the drop with applicator stick or glass rod, being careful not to contact the surface of the slide.
6. Incubate the slide in a humid chamber at room temperature for at least 25 minutes. If using a 37° C incubator, incubate for 25 minutes. Longer incubation periods are OK.
7. Rinse the slide free of antibody reagent by adding 50 – 100 uL SureRinse™ wash buffer and leave for 1 minute. Tilt slide, long edge down, and absorb excess fluid with absorbent material placed at the edge of the slide well. Do not touch the surface of the well slide or disturb the sample.
8. Non-specific background fluorescence may be reduced, and a reddish background added to enhance contrast, by the use of BlockOut™ counterstain at this stage. Apply 1 drop of counterstain per well. Incubate for 1 minute at room temperature.
9. Rinse the slide free of counterstain by adding 50 – 100 uL SureRinse™ wash buffer and leave for 1 minute. Tilt slide, long edge down, and absorb excess fluid with absorbent material placed at the edge of the slide well. Do not touch the surface of the well slide or disturb the sample.
10. The slide should be partially-to-completely air dried on a slant and then mounted with one drop (~45 uL) of No-Fade™ mounting medium. Apply cover glass and view.

## Other Information, Tips & Troubleshooting

1. Test Time: Approximately 35 – 40 minutes after the sample is dried to the well slide and without methanol fixation step. (Approximately 1.0 hr when performing methanol fixation.)
2. ST102R, Giardia-a-Glo™ reagent will stain both viable (live) and non-viable (dead) cells. It will stain cysts and oocysts preserved by gamma irradiation or suspended in formalin.
3. When making a positive control slide using PC101, mix the contents of the vial prior to use. Vortex the vial for 20 seconds immediately before use. Note: The number of organisms in PC101 is not exact and should not be used for sample recovery estimation.
4. Prepared slides (mounted with M101, No-Fade™ mounting medium) may be kept in a refrigerator/protected from light and viewed repeatedly for 6 months or longer. DAPI staining may fade.
5. Steps 3 & 4 can be performed after steps 5 & 6, that is, DAPI may be applied to the sample well either before staining with Giardia-a-Glo™ or after.
6. If DAPI staining appears faint, the reaction time may be increased from 1 minute to 4 minutes. Another option is to increase the concentration to 1 ug/mL. To dilute DAPI to 1 ug/mL, add 2.5 uL D101 to 5 mL PBS or 25 uL DAPI to 50 mL PBS. If DAPI staining continues to be faint, the concentration can be increased further to 2 ug/mL. To dilute to 2 ug/mL, add 5 uL D101 to 5 mL PBS or 50uL D101 to 50 mL of PBS.
- 7.
8. One resource available to help distinguish between *Giardia* cysts, *Cryptosporidium* oocysts and possible cross-reactors can be found on the US EPA website. The US EPA has developed training modules for the Long Term 2 (LT2) Enhanced Surface Water treatment Rule. These training modules were developed to assist analysts in the detection and identification of *Giardia* and *Cryptosporidium*. They can be found at: [www.epa.gov/safewater/lt2/training/index.html](http://www.epa.gov/safewater/lt2/training/index.html).

---

For assistance, technical questions, or to inquire about other Waterborne™, Inc. products, please call, FAX, or e-mail us. Also, please visit our website at [www.waterborneinc.com](http://www.waterborneinc.com).