

AOS Suprapatellar Instruments


Assemble instruments as shown.



Use with Extended Insertion Guide (1255) and Extended Bolt (1256).




Track instrument assembly over guide pin to entry point.




Secure location by fixing handle to femur with pins if desired.


Unscrew and remove Pin Guide.




Track nail over guide wire and thru cannula, which will flex when passing proximal bend of nail.



Positioning nail and insert screws using standard procedure.





Slide release button as indicated by arrow
to release disposable cannula

